

N.I.S.T. CERTIFICATION

ANY SPAN Instrument can be tested, serialized and certified traceable to N.I.S.T. accuracy standards.

We offer General Certificates of Accuracy and Calibration Certificates with Test Points.

Please ask one of our product specialists for details.

SPAN CALIBRATION CERTIFICATE

Description LFS212-30-PSI-G-OCEANEERING
Catalog No. 4102132 Serial No. S27697
Operating Range 0 TO 30 PSI Accuracy \pm #####

CALIBRATION DATA

INCREASING PRESSURE			DECREASING PRESSURE		
Applied press.	Indicated Press.	Difference	Applied press.	Indicated Press.	Difference
5		-5	25		-25
10		-10	20		-20
15		-15	15		-15
20		-20	10		-10
25		-25	5		-5
30		-30			

Allowable Operating Tolerance \pm 0.3 PSI

Span certifies that this instrument has been tested and calibrated with test standards traceable to N.I.S.T. as specified in American National Standards ASME B40.1-1998, PP6.1 And test procedures as specified in ASM B40.1-1998, PP6.2

Calibrated in upright Position Temperature 72 DEG
Date of Certification 7/11/2012 Inspector 

THUEMLING INSTRUMENT GROUP

MANUFACTURER OF SPAN GAUGES
P O BOX 1625 WAUKESHA, WI 53187-1625 PHONE (262) 547-1789
1225 PEARL STREET WAUKESHA WI 53186-4887 FAX (262) 547-6493